

# Medical Suction Filters

for compliance with safety standards

## Comprehensive Range

- | Virus, bacteria and odour filters
- | Medela and conical connections
- | Disposable

## High-Performance

- | Dedicated, high-performance medical membranes
- | Lab-tested high bacteria retention grades
- | Low air resistance due to the large surface area



Precious life,  
progressive care

**Medela offers a vast portfolio of disposable filters that are seamlessly compatible with all Medela healthcare devices. They enable you to custom-configure your suction system to suit a variety of hygienic requirements.**

- Bacteria filters provide the first, fundamental level of hygienic protection. A basic requirement if your system does not include an embedded bacterial filter.
- Overflow filters offer an essential safety mechanism to prevent fluids from entering the pump or central vacuum system. Two redundant overflow mechanisms should be present in any aspiration system.
- Odour membranes effectively trap gases and odours, especially useful in containers used over multiple hours.
- To ensure safety for patients and hospital staff, the pump filter exchange frequency should be performed according to the instructions for use and in-line with applicable local infection control policy.
- Filtering capacity (Bacteria filter) 99.985% for particles larger than 0.3 µm.
- Filtering capacity (Virus filter) 99.99984% for particles larger than 0.03 µm.



Medela Connection



Conical Connection

## Filter Overview

Type of filter	Medela/Medela connections	Conic/conic connections	Recommended application	For following collection systems
Bacteria	Ref. 077.0573		Basic hygienic protection	Reusable jars
Virus	Ref. 101035265		Basic hygienic protection	Reusable jars
Bacteria/Overflow	Ref. 077.0571	Ref. 077.0572	Second stage overflow protection	Disposable liners, reusable jars
Virus/Overflow	Ref. 101035263	Ref. 101035264	Second stage overflow protection	Disposable liners, reusable jars
Odour/Bacteria	Ref. 077.0575		Reduction of odour	Disposable liners, reusable jars

## Filter Exchange

- In case of overflow or clogging, or
- Weekly with single-patient use, or
- Daily with multiple patient use.

To detect a clogged filter, perform a filter test before each patient use. Comply with in-house directives on hygiene.

Medical Vacuum Technology for Healthcare Professionals

**Please contact us or your local Medela representative for details.**


 Medela AG  
 Lättichstrasse 4b  
 6340 Baar  
 Switzerland  
[www.medela.com](http://www.medela.com)

Local contact

